



EU Declaration of Conformity as per MDR 2017/745

<i>Manufacturer:</i>	PRODUITS DENTAIRE SA Rue des Bosquets 18 1800 Vevey Switzerland
<i>CHRN</i>	CHRN-MF-20000511
<i>SRN</i>	CH-MF-000023147

<i>European Authorized Representative</i>	PD Dental EU Thonon-les-Bains, 74200 France
<i>European Authorized Representative SRN</i>	FR-AR-000017144

We declare under our sole responsibility that the medical device of the following Generic Device Group:

Plastic Pipettes

<i>Reference:</i>	See attached table
<i>Risk class:</i>	I
<i>As per rule:</i>	5 Annex VIII, MDR 2017/745
<i>Intended Purpose:</i>	Device used for delivering small quantities of liquid into the root canal or pulp chamber during endodontic or restorative treatment
<i>Trade name:</i>	Plastic Pipettes
<i>Basic UDI-DI:</i>	764015916PIPETTESB4

Meets all the requirements of the medical device regulation **MDR 2017/745**

Conformity assessment procedure: MDR 2017/745 Article 52, Annex II and III and ODim July 2020.

Quality & Regulatory Affairs Engineer

Silvia Contini

Vevey, 2022-04-22

President

Nicolas Gehrig

Vevey, 2022-04-22



EU Declaration of Conformity as per MDR 2017/745

Plastic Pipettes

<i>Reference</i>	<i>Designation</i>
11750	PLASTIC PIPETTES 0.3mL ASSORTED (5)
11752	PLASTIC PIPETTES 1mL ASSORTED (5)
11754	PLASTIC PIPETTES 0.3mL ASSORTED (25)
11757	PLASTIC PIPETTES 1mL ASSORTED (25)