

EU Declaration of Conformity as per MDR 2017/745

Manufacturer:

PRODUITS DENTAIRES SA

Rue des Bosquets 18

1800 Vevey Switzerland

We declare under our sole responsibility that the medical device:

PI	ast	ic F	ipe	ttes
----	-----	------	-----	------

Reference:

See attached table

Risk class:

1

As per rule:

5 Annex VIII, MDR 2017/745

Intended Purpose:

Device used for delivering small quantities of liquid into the root canal or pulp

chamber during endodontic or restorative treatment

Basic UDI-DI:

07640159161462

Meets all the requirements of the medical device regulation MDR 2017/745

Conformity assessment procedure:

MDR 2017/745 Article 19 and Annex IV

Registered with:

Swissmedic Hallerstrasse 7 3012 Bern Switzerland

Quality & Regulatory Affairs Director

Chloé Fattebert

Vevey, 2021-05-20



EU Declaration of Conformity as per MDR 2017/745

Plastic Pipettes

Reference	Designation	
11750	PLASTIC PIPETTES 0.3mL ASSORTED (5)	
11752	PLASTIC PIPETTES 1mL ASSORTED (5)	
11754	PLASTIC PIPETTES 0.3mL ASSORTED (25)	

F 25 02 - DeCo MDR - Pipettes - v01 2 / 2